



Fiscal Year 2020

CERTIFICATE OF FDA REGISTRATION

This certifies that:

HENAN ZHONGJIAN MEDICAL EQUIPMENT CO., LTD

Dingluan Town Pioneer Park

Changyuan, Henan, 453400, CHINA

DUNS Number:

has completed the FDA Establishment Registration and Device Listing with the US Food & Drug Administration, through UCL-REG SERVICE INC.

Owner/Operator Number: 10065214

Listing No.

D380918

Product Code:

MSH

Device Name:

Respirator, surgical

N95 protective mask

UCL-REG SERVICE INC. will confirm that such registration remains effective upon request and presentation of this certificate until the end of the calendar year stated above, unless said registration is terminated after issuance of this certificate. UCL-REG SERVICE INC. makes no other representations or warranties, nor does this certificate make any representations or warranties to any person or entity other than the named certificate holder, for whose sole benefit it is issued. This certificate does not denote endorsement or approval of the certificate-holder's device or establishment by the U.S. Food and Drug Administration. UCL-REG SERVICE INC. assumes no liability to any person or entity in connection with the foregoing.

Pursuant to 21 CFR 807.39, "Registration of a device establishment or assignment of a registration number does not in any way denote approval of the establishment or its products. Any representation that creates an impression of official approval because of registration or possession of a registration number is misleading and constitutes misbranding."

The U.S. Food and Drug Administration does not issue a certificate of registration, nor does the U.S. Food and Drug Administration recognize a certificate of registration, UCL-REG SERVICE INC. is not affiliated with the U.S. Food and Drug Administration.



UCL-REGSERVICE INC.

602 ROCKWOOD ROAD, WILMINGTON,

NEW CASTLE DE 19802 USA

For and on behalf of
UCL-REG SERVICE INC.

Authorized Signature(s)

Cert. No.: M20717

Issued Date: 28 March 2020

Expiration Date: 31 December 2020