



Intco Medical(HK) Co., Ltd.

Add:No.18 Qingtian Road Qilu Chemical Industrial

Zibo,Shandong,China

Web:www.intcomedical.com

Synmax Vinyl Examination Disposable Gloves Product Specification

Type	Synmax Vinyl Examination Disposable Gloves, Powder free, Non Sterile, Ambidextrous				
Grade	Examination	Surface	Smooth	Glove Length(mm)	230mm
Material	Poly-Vinyl Chloride	Packing	100pcs/box,10boxes/ctn	Weight(g)	5.2g (Medium Size)
Cuff	Beaded			Powder Level	≤2mg/glove

Product	Color	Size	Weight	Physical Dimension					Physical Property		Appearance AQL	
				Length (mm)	Palm Width (mm)	Cuff Thickness (mm)+-0.03	Finger Thickness (mm)+-0.03	Palm Thickness (mm)+-0.03	Tensile Strength(MPA)			Water Test AQL
									Before Aging	After Aging		
Synmax Vinyl Examination Disposable Gloves	Blue	S	4.6g+-0.3	230	85±5	0.04	0.09	0.07	13	11	AQL1.5	Critical 0 Major AQL1.6 Minor 2.6 Major + Minor 4.0%
		M	5.2g+-0.3	230	95±5	0.04	0.09	0.07	13	11	AQL1.5	
		L	5.4g+-0.3	230	105±5	0.04	0.09	0.07	13	11	AQL1.5	
		XL	5.7g+-0.3	230	115±5	0.04	0.09	0.07	13	11	AQL1.5	

No.	sort	Critical defects	Major defects	Minor defects
1	torn	big tearing	√	-
		scratch	≥10mm	<10mm
		visible hole	≥10mm	<10mm
2	dirt	greasy dirt	≥10mm,serious	5-10mm,visible
		yellow dirt	√	-
3	black/PU spots,impurities	uncountable	≥3mm or 0.5-3mm,>3/side	0.5-3mm,≤3/side
4	beading	broken;torn;none-beading	impurity,black,bright,PU 3mm;inward≥10mm;minuteness(≤0.5mm)	bright PU 0.5-3mm; inward<10mm;thin 0.5-0.8
5	sticky	-	sticky cuff severely	slightly
6	Extreme color variation	entirely,severely	finger,cuff,severely	cuff,slightly,unconspicuous
7	Bad smell	-	severely irritating smell	slight odor
8	Embedded foreign particles	insect,hair,etc.	-	-

We has established the Quality management system in according with ISO9001 and ISO 13485 standards

The certification which we have is 510K ,CE certification ,ISO9001 ISO13485,NSF,CFIA

Sampling Procedures according to the standard of ISO2859,ASTM D5250,EN455 ,Inspection Level for physical dimensions and force /elongation at break is according to S-2 ,water test according to G-1